


CF 7

Pentode

Hauptinstrument
Bereich



● **6mA**

● A 200 Volt

Sollwert: 3mA

● ● ●

● U_{III} 100 Volt

G₂ ●

● U_I 2 Volt

G₁ ●

● ●

G₃ K

F ●

Heizung: 13 Volt/200 mA

● F

Prüfkarte für RPM 370
Neuberger, München 25