



# ECL 86

## Pentode Triode



Noval B9A



$J_A$  - Bereich

A

$J_A$  **36** mA

250 V

60

$U_{III}$

$G_2$

$U_I$

250 V

$G_1$

$G_{1,T}$

$K,T$

K

$A_T$

7V

F

F

Heizung 6,3 Volt

Meßkarte für NEUBERGER Röhrenmeßgeräte

Meßspannungen

Heizung	F	6,3	V
$U_I$	$G_1$	7	-V
$U_{II}$			-V
$U_{III}$	$G_2$	250	V
$U_{IV}$			V
A		250	V

Richtströme

$J_A$		36	mA
$J_{G_2}$			mA
S			mA/V