

Betriebsdaten:
 Typical operation:
 Exemple d'utilisation:

U_f . . . 6,3 V
 I_f . . . 0,2 A
 U_a . . +250 V=
 U_g . . . 0
 R_a . . . 1 M Ω
 U_L . . +250 V=
 I_L . . . 0,75 mA



Prüfdaten:
 Test operation:
 Vérification:

U_f . . . 6,3 V \sim
 U_a . . +100 V=
 U_{g1} . 0 & -2 V=
 U_L ($\neq U_{g2}$) .
 . . . +150 V=

In Stellung 12 und 13
 muß Abstimmfeld leuchten,
 sonst ist die Röhre
 unbrauchbar.
In position 12 and 13
 the tuning indicator has to be
 shining, but he is bad.
Il faut, que le tube indi-
cateur luit en Position 12 et
13, autrement il est inutile.

313

Patent-Röhrenprüfer Modell W 19

MAX FUNKE, Adenau-Eifel 7. 54.


EM 4

Abstimmanzeiger / Tuning indicator / Tube indicateur.

Unbrauchbar
 mauvais / Bad

?

G u t
 Good / Bon




2,5 mA

Grenzdaten:

Ratings / Conditions maxima:

U_a +300 V=
 U_L max +300 V=

U_L min +90 V=
 R_{fk} . . 20 k Ω
 R_g . . . 3 M Ω
 U_{fk} . . 100 V



In Stellung 13
 auf Steuerwirkung
 prüfen
In position 13 is to
prove the control grid.
En position 13 vérifier
l'action de commande de la
grille.