


# QQE 03/12


6360

## Tetrode II

### Tetrode I



Noval B9A



# $J_A$ 28 mA

$A_{II}$

$G_2$  200 V

$U_{III}$

$G_2$

$U_{II}$

$G_{III}$

150 V

$G_{II}$

$A_I$

$U_I$

10 V

K

85 V

F

F

### Heizung 12,6 Volt

Meßkarte für NEUBERGER Röhrenmeßgeräte

#### Meßspannungen

Heizung	F	12,6	V
$U_I$	$G_I$	85	-V
$U_{II}$	$G_{III}$	10	-V
$U_{III}$	$G_2$	150	V
$U_{IV}$			V
A		200	V

#### Richtströme

$J_A$		28	mA
$J_{G_2}$			mA
S			mA/V