



# QQE 03/20

# Tetrode I

RS 1019

6252



# Tetrode II

# $J_A$ 40 mA

$A_1$

$J_A$  - Bereich



400 V

120

$U_{II}$

$G_{III}$



$U_{III}$

$G_2$



$U_I$

85 V



250 V

$G_{II}$



K



FM

15 V

F



F



## Heizung 6,3 Volt

Meßkarte für NEUBERGER Röhrenmeßgeräte

### Meßspannungen

Heizung	F	6,3	V
$U_I$	$G_{II}$	15	-V
$U_{II}$	$G_{III}$	85	-V
$U_{III}$	$G_2$	250	V
$U_{IV}$			V
A		400	V

### Richtströme

$J_A$	40	mA
$J_{G_2}$		mA
S		mA/V