



**Betriebsdaten:**  
 Typical operation:  
 Exemple d'utilisation:

	1304	1404
$U_f$ . . . . 4	. . . . 4 V	
$I_f$   . . . . 1.1	. . . . 1.3 A	
$U_a$ . max 500	. 800 V <sub>eff</sub>	
$I_a$ . max 100	. 100 mA	

**Prüfdaten:**  
 Test operation:  
 Vérification:

$U_f$  . . . . 4 V<sub>~</sub>  
 $U_a$  . . . . 60 V<sub>~</sub>

Patent-Röhrenprüfer Modell W19

MAX FUNKE, Adenau-Eifel 9. 54.

Telefunken	Valvo	Tungsram	Philips
RGN 1304	G 495	V 495	505
RGN 1404	G 1404	G 4205	V 4200
			1832

Einweg-Gleichrichter / Half wave rectifier / Redresseur monoplaque.

Unbrauchbar / mauvais / Bad      ?      **G u t** / Good / Bon     

